



NQAC

Nestlé Quality Assurance Center  
Dublin

# Technical Datasheet

**Analysis Name:** Total Iodine Analysis by ICP-MS

**Method Number:** LI-00.849

**Scope of Application:** This instruction describes the quantification of total iodine by ICP-MS in foods, beverages (finished, concentrates, and powders), health products, pet foods, gummy-based vitamins, and raw materials such as premixes, food grade oils, salts, and tastemakers.

**Description:** Iodine is extracted from the test portion of the sample with a strong alkaline reagent in a closed vessel microwave digestion system. Digested samples are ionized through inductively coupled argon plasma. The ions of iodine are extracted from the plasma, separated in the mass spectrometer, and determined using a pulse counting detector system.

**Sample Weight Required:** 50 g

**Method Reference:** CEN EN 15111-2007. Foodstuffs. Determination of iodine by ICP-MS  
LMBG, Analysis of foods. Determination of iodine in dietetic foods by ICP-MS (MS with inductively coupled plasma).  
Amtliche-Sammlung-von-Untersuchungsverfahren-nach-Paragraph-35-LMBG; L 49.00-6, (1998).

**Analytical Platform:** ICP-MS

**Special Information:** Include Certificate of Analysis or estimated levels for premix samples.

Analyte Reported	Alias	Unit of Measure	Limit of Quantification	Reproducibility
Iodine		µg/100 g	3.0	25%